

X-636-65-262

FACILITY FORM 602

N 66-11205  
15  
C  
NASA CR OR TMX OR AD NUMBER

(THRU)

(CODE)

(CATEGORY)

NASA TM X-55309

# SOLID STATE AMPERE HOUR INTEGRATOR

BY  
JOHN PAULKOVICH

GPO PRICE \$

CFSTI PRICE(S) \$

Hard copy (HC) 1.00

Microfiche (MF) .50

ff 653 July 65

JULY 1965



GODDARD SPACE FLIGHT CENTER  
GREENBELT, MARYLAND

X-636-65-262

**SOLID STATE AMPERE  
HOUR INTEGRATOR**

by  
**John Paulkovich**

**July 1965**

**Goddard Space Flight Center  
Greenbelt, Maryland**

## ILLUSTRATIONS

<u>Figure</u>		<u>Page</u>
1.	Block Diagram of Ampere Hour Integrator . . . . .	5
2.	Current Transformer Circuit . . . . .	5
3.	Current to Voltage Converter . . . . .	6
4.	Current to Frequency Converter . . . . .	6
5.	Schematic of Ampere Hour Integrator . . . . .	7
6.	Integrator Power Supply . . . . .	8
7.	Count Rate vs Current . . . . .	9
8.	Ampere Hour Integrator . . . . .	10
9.	Internal View of the Ampere Hour Integrator . . . . .	11
10.	Ampere Hour Integrator and Print Out Counter Mounted in a 19 inch Panel . . . . .	12

## SOLID STATE AMPERE HOUR INTEGRATOR

By

John Paulkovich

One of the problems in evaluating satellite battery packs is to determine the ampere hour efficiency accurately. Normally this is accomplished by recording the current rate on a strip chart recorder. The data is then processed by summation of the area under the curve to determine the ampere hours accumulated. This has always been a slow and cumbersome task where the accuracy depended upon the strip chart recorder and was further aggravated by approximations by the personnel performing the function.

This paper describes a solid ampere hour integrator designed for evaluation of space satellite batteries. The integrator has to meet the following requirements:

- The unit is to accumulate ampere hour capacities over the range of .01 ampere rate to 20 ampere rate (in two ranges).
- The accuracy to be better than  $\pm 1$  percent.
- The accumulated ampere hours to be read directly.
- An output pulse to be available for operation of an external print-out counter.

Since ampere hours represent the product of current and time then the object is to convert this product into units that may be accumulated with time or to put it another way to convert the ampere rate to pulse rate and accumulate these pulses.

Basically the ampere hour integrator is a current to frequency converter. Figure 1 illustrates a power supply, battery, current to frequency converter, and a counter. The current to frequency converter has an output pulse rate directly proportional to the battery charge current. The counter accumulates the total number of pulses and displays this total thereby indicating the accumulated ampere hours.

The ampere hour integrator discussed here is composed of three basic units:

1. Current transformer.
2. Current to frequency converter.
3. Accumulator.

## CURRENT TRANSFORMER

Figure 2 illustrates a basic circuit of the current transformer used. Circuit operation is as follows:

Assuming  $I_1$  is the current to be sensed and is zero at this time, and further assuming  $Q_1$  is on, then the transformer is reset to negative saturation. The degree of saturation is controlled by the limiting resistor  $R_1$ .  $I_2$  will be zero during this time interval. Turning on  $Q_2$  will cause the core to go toward positive saturation. During this period a small magnetizing current will be indicated by  $I_2$ . (It might be noted that  $N_2$  does not saturate). This current ( $I_2$ ) will be equal to the magnetizing current when  $I_1$  is zero. Now suppose a current  $I_1$  is applied to  $N_3$  in such a direction as to add to  $I_m$ , then  $I_2$  will be composed of the initial magnetizing current plus the reflected current of  $I_1$  which is

$$I_2 = I_m + (N_3/N_2) I_1$$

This current will be directly proportional to  $I_1 + I_m$ .  $I_m$  is an undesirable component which cannot be eliminated but whose effects can be minimized.

Figure 3 illustrates a circuit whereby this may be accomplished.

Advantage is taken of the  $B_r$  and  $B_m$  characteristics of the transformer core in this circuit.  $B_m$  less  $B_r$  actually represents a stored inductive charge in the transformer  $T_1$ . As a result when  $Q_1$  turns "off" an undesirable voltage spike is developed across the windings of the polarity indicated in Figure 3. This voltage is in excess of  $E_{cc1}$  and is clamped at  $Q_2$  by diode  $D_1$ . This in turn causes the base emitter of  $Q_3$  to bias in the "off" direction. If capacitor  $C_1$  stores this energy then  $Q_3$  will maintain a cut-off bias when  $I_1 = 0$ .  $C_1$  is now capable of supplying the magnetizing current necessary for  $N_2$ .  $R_2$  is adjusted for an equilibrium balance to drain off the excess stored charges of  $C_1$ . Now a current  $I_1$  will be reflected to  $N_2$  and will be the major collector current of  $Q_3$ . This in turn will induce a voltage on  $R_3$  directly proportional to  $I_1$  and will be

$$E_{R_3} = \frac{a(N_3/N_2) I_1}{2}$$

where  $E_{R_3}$  = average DC voltage on  $R_3$

$a$  = alpha of  $Q_3$

$N_3/N_2$  = turns ratio of the transformer

$I_1$  = current being sensed

$R_3$  = DC resistance of  $R_3$

At this point the circuit is a "current to voltage converter". Substituting a uni-junction transistor circuit for  $R_3$  will make this a "current to frequency converter."

For the circuit illustrated in Figure 4, the rate of charge of  $C_2$  will be directly proportional to  $I_1$  and the output pulse rate ( $F_0$ ) will also be directly proportional to  $I_1$ . Either a positive going or negative going output pulse is available at  $Q_4$ . This in effect comprises the basic circuit of the current to frequency converter portion of the ampere hour integrator.

Figures 5 and 6 illustrate the complete schematic of the ampere hour integrator. Referring to Figure 5,  $Q_1$  and  $Q_4$  and the associated circuitry compose a Royer oscillator as a clock reference and also as the driver for  $Q_2$  and  $Q_3$  of the current sensing circuit. Since the current sensing is uni-directional, two relays,  $K_1$  and  $K_2$  reverse the sensing direction upon closure of the external contacts of  $J_3$ . The relays reverse the low current primary windings of  $T_2$  rather than the high current sensing winding. Uni-junction transistor  $Q_6$  converts the ampere rate to pulse rate. These pulses are fed to  $Q_8$  via resistor  $R_{10}$ .  $Q_7$  and  $Q_8$  compose a single shot multivibrator whose "on" time is determined by  $C_6$  and  $R_{15}$ . The collector load of  $Q_8$  is either a relay  $K_3$  (for operating an external counter) or the self contained counter of the unit.

Two current ranges are incorporated in the unit. The high current range is composed of  $L_1$  and one turn winding on  $T_2$ .  $L_1$  is necessary to reduce the effects of external circuit impedance upon the current sensor. This range is from 100 milli-ampere rate to a 25 ampere rate. When the low range is used, the total turns (1 + 9) and both inductors  $L_1$  and  $L_2$  are connected in series. This reduces the range by a factor of 10 or 10 milli-ampere rate to 2.5 ampere rate, thus allowing sufficient overlap of the two ranges.

Figure 7 shows a graphical plot of the counting rate vs sensing amperes.

This is a plot of the high current range. The low current range is an exact replica of this plot if the sensing current scale is multiplied by 1/10. The circuit exhibits excellent linearity from .10 ampere rate to 20 amperes with slight non-linearity below and above this range.

## CALIBRATION

Calibration of the instrument is rather simple and is accomplished on the low range.  $C_7$ ,  $C_8$ , and  $C_9$  are selected for the desired count at two amperes

sensing current (Figure 5). The low end is then checked at a 20 ma. rate. The low end is adjusted by  $R_8$ . Once  $C_7$ ,  $C_8$ , and  $C_9$  have been selected, final trim adjustment at the high end is accomplished at the power supply by  $R_4$  (Figure 6). Since a change in range is essentially a change in ampere turns, both ranges are calibrated simultaneously.

Two regulated power supplies are incorporated (Figure 6), 30 volt regulated supply for the current to frequency converter circuit and a 12 volt simple regulator for the counter and relays. This eliminates interaction of the high current of the counter from the current sensing circuits.

The low range accuracy is better than  $\pm 1$  percent from 10 ma. rate to a 2 ampere rate and within  $\pm 2$  percent up to 2.5 amperes. The same applies to the high range which is 100 ma. rate to 20 ampere rate ( $\pm 1\%$ ) and to 25 amperes ( $\pm 2\%$ ).

Figures 8, 9, and 10 are pictures of the Ampere Hour Integrator. Figure 8 shows the Ampere Hour Integrator in a 8 x 8 x 5 case. Figure 9 shows the internal construction and Figure 10 shows the Ampere Hour Integrator mounted in a 19 inch rack panel along with a "print out counter".

The addition of the "print out counter" increases the versatility of the unit. As mentioned previously the sensing direction of the Ampere Hour Integrator can be reversed upon closure of external contacts. "Print out" and "reset to zero" of the print out counter can also be controlled from a remote point, thus permitting remote programming of the Ampere Hour Integrator—Print Out Counter combination.

The use of the ampere hour integrator simplified battery evaluation and also permitted more tests to be conducted simultaneously without materially increasing the work load on the personnel. Several units have been in continuous operation for over a year and indicate no significant change in calibration or performance.




Figure 1 - Block Diagram of Ampere Hour Integrator




Figure 2 - Current Transformer Circuit






Figure 3 - Current to Voltage Converter





Figure 4 - Current to Frequency Converter






Figure 6 - Integrator Power Supply




Figure 7 - Count Rate vs Current




Figure 8 - Ampere Hour Integrator




Figure 9 - Internal View - Ampere Hour Integrator




Figure 10 - Ampere Hour Integrator and Print Out Counter in a 19 inch Panel